

Settore **LABORATORIO**


Laboratory balances




laboratory
balances



industrial
scales



instruments
for paint



oenological
equipments



digital
tensiometer



calibrated
masses



laboratory electronic balances & oenological equipments



customer care

info@gibertini.com

Tel. (+39) 02 3541434

GIBERTINI ELETTRONICA SRL

Via Bellini, 37
20026 Novate Milanese
(Milano) - ITALY
Tel. (+39) 02 3541434
Fax (+39) 02 3541438
sales@gibertini.com

www.gibertini.com

ISO 9001:2015



LAT N° 094 RMP N° 094

Il Centro ACCREDIA LAT n. 094 è accreditato per tarature di bilance e di microdosatori.
Il Centro ACCREDIA RMP n. 094 è accreditato come produttore di materiali di riferimento (soluzioni idroalcoliche). Per i dettagli consultare le tabelle di accreditamento su www.accredia.it.
ACCREDIA LAT Center n. 094 is accredited for balances and microdosimeters calibration.
ACCREDIA RMP Center n. 094 is accredited as reference material producer (hydroalcoholic solutions).
For details consult the accreditation tables on www.accredia.it.

MAGNETIC COMPENSATION SEMIMICRO ANALYTICAL BALANCE DUAL RANGE

Mod. E50S/3

Functions

- Automatic autocalibration with built-in calibration weight
- Prepared for weighings under the balance plane
- Stable weight determination signal
- Acid resistant membrane keyboards. Parameters configurable by menu: reading in g (grams), lb (pound), oz (ounce), ct (carats), pcs (pieces), % (percentage)

Technical specifications

- Reading by LCD display
- Bubble and feet for levelling
- Pan dimension: diam. 80 mm
- Data output: RS232 I/O serial output
- Operating temp.: 10° ÷ 35°C
- Power supply unit: 230V (-15% / +20%) 50Hz (110V on request)
- Absorption: 12 VA
- Total dimension: W 210, D 340, H 320 mm
- Weighing chamber dim.: W 180, D 170, H 220 mm
- Net weight: 6,6 kg

On request

- Class E2 masses
- Solids density measuring devices
- Printer

Mode	Capacity g	Readability mg	Linearity mg	Repeatability mg	Response time
Semimicro	100	0,01	± 0,03	± 0,03	6/8 sec.
Macro	225	0,1	± 0,2	≤ 0,05	3/5 sec.

Subject to technical changes without notice.

All our instruments are to be used exclusively for scientific purpose and/or for internal control only. They must not be used in environments with danger of explosions.

The use of our balances is forbidden in the cases provided by art. 1, point 2, letter a) of directive 2014/31/UE except for those models approved with CE mark (legal metrology).

Electronic balances are sensible to the variation of the acceleration of gravity. They must be calibrated on the working site (2014/31/UE).

All our instruments comply with directive 2004/108 UE (electromagnetic compatibility)