



# ***Circulating fluid temperature controller THERMOCHILLER***



*Subject to technical changes without notice.*

*All our instruments are to be used exclusively for scientific purpose and/or for internal control only. They must not be used in environments with danger of explosions. The use of our balances is forbidden in the cases provided by art. 1, point 2, letter a) of directive 2014/31/UE except for those models approved with CE mark (legal metrology). Electronic balances are sensible to the variation of the acceleration of gravity. They must be calibrated on the working site (2014/31/UE). All our instruments comply with directive 2014/30 UE (electromagnetic compatibility)*



# ***Circulating fluid temperature controller THERMOCHILLER***

*Optimized configuration for use with Gibertini Oenological  
Electronic Distillers model SuperDEE and DEE*

- Temperature range from 5°C to 40°C
  - Temperature stability  $\pm 0,1^{\circ}\text{C}$
  - Electronic temperature control and temperature reading on digital display
  - Tank capacity 5 liters
  - Integrated pump for recirculating of the cooled liquid
  - Mounted on pivoting wheels for easy movement
  - Power supply 230V 1700W – 50/60 Hz, other voltages on request
  - Dimensions (W x D x H) 38 x 50 x 62 cm
  - Net weight 43 Kg
  - Ambient temperature 5° to 40°C, humidity 30 to 70% (not condensing)
- 
- Space saving and quiet operation (60 dB)
  - Low tank level detection function
  - Power failure auto-restart function
  - Anti-freezing operation function
  - Self-diagnosis function and check display
  - Easy and tool-less filter maintenance
- 
- Special executions available on request

Subject to technical changes without notice.

All our instruments are to be used exclusively for scientific purpose and/or for internal control only. They must not be used in environments with danger of explosions.  
The use of our balances is forbidden in the cases provided by art. 1, point 2, letter a) of directive 2014/31/UE except for those models approved with CE mark (legal metrology).  
Electronic balances are sensible to the variation of the acceleration of gravity. They must be calibrated on the working site (2014/31/UE).  
All our instruments comply with directive 2014/30 UE (electromagnetic compatibility)

June 2019